



AMERETAT SHIMI

Specification

Water



Tests	Standards
Spec. conductance at 25 °C (at the time of manufacturing)	$< 1\mu S/cm$
Identification	<i>Complies</i>
Gradient grade	
at 210 nm	$\leq 5 mAU$
at 254 nm	$\leq 0.5 mAU$
basic absorption 210 nm	$< 20 mAU$
Fluorescence	
as quin. at 254 nm	$\leq 1.0 ppb$
as quin. at 365 nm	$\leq 0.5 ppb$